

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JAN 26 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130303</u> ✓
Name <u>B.H. DUDLEY</u>		ID# <u>3532</u> Date <u>1.22.10</u> ✓
A	Agency <u>Wasilla P.O.</u>	Phone # _____
Instrument Location <u>Wasilla, AK</u>		
Alco S/N <u>75509</u> ✓ Target Value <u>.079</u> High Pressure <u>400</u>		
B	Alco Test Value Average <u>.078</u> ✓ 1 <sup>st</sup> Alco <u>.078</u> 2 <sup>nd</sup> Alco <u>.081</u> <i>(4-26-10 C.M.)</i>	
Signature <u>B.H. Dudley</u> <i>CB</i> 1/26/10		
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

JANUARY 22, 2010

OPERATOR'S NAME:

SUPERVISOR

OPERATOR'S NUMBER: 3532

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: WAS1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: 75509

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.54 in

ALCO TARGET .078

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.54 in .081

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 29.54 in .081

BLANK TEST .000

16:32

16:33

16:33

16:34

16:34

16:35

16:36

16:36

16:37

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

JANUARY 22, 2010

TIME 16:29

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.54 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~Δ